1 INTRODUCTION

The Marine Biological Laboratory (MBL) is committed to providing a healthy and safe learning, teaching, research, and work environment. The Biosafety Manual is part of the MBL's biological safety program designed to minimize exposure and mitigate the risk associated with handling potentially infectious biological materials.

This manual serves as a resource document for working safely in research and teaching laboratories at MBL in order to:

(a) Inform and educate laboratory personnel and students on biosafety practices and procedures.
(b) Protect the MBL community (staff, students, and visitors) and the general public from exposure to potentially infectious agents and recombinant DNA.
(c) Prevent environmental contamination.
(d) Provide an environment for high quality learning, teaching, and research while maintaining a safe workplace.
(e) Comply with all pertinent federal, state, and local regulations and guidelines.

The Biosafety Manual provides institutional-wide safety guidelines for working with biological hazards. Biosafety practices and procedures in all research and teaching laboratories at the MBL should co

1.3 Biosafety Manual Review

The Biosafety Manual shall be reviewed and updated, as necessary, to reflect new or revised regulations, or modified tasks and procedures. Submit any suggestions and comments regarding this manual to the Biosafety Officer at 508-289-7424 or safety@mbl.edu.

2 ROLES AND RESPONSIBILITIES

The specific responsibilities for implementation of the biosafety program at MBL are outlined below:

2.1 Institutional Biosafety Committee (IBC)

The Institutional Biosafety Committee (IBC) is an MBL Standing Committee responsible for reviewing research activities using recombinant DNA (rDNA), infectious agents, Select Agents, non-native species, and projects with blood products. In addition the IBC is involved with overseeing the safe conduct of research, and assessing biocontainment levels. The Committee provides oversight of the biosafety program to ensure research is conducted in compliance with federal, state and local regulations and MBL policies. The "NIH Guidelines for Use of Recombinant or Synthetic Nucleic Acid Molecule**s** " require that all procedures involving the use of rDNA undergo review by the IBC.

https://osp.od.nih.gov/biotechnology/nih-guidelines/

For more information related to MBL's Institutional Biosafety Committee (IBC), please refer to the IBC Policy & Procedures Manual. The manual can be found at:

https://www.mbl.edu/research/health-safety/biological-safety/biosafety-committee

2.2 Department Chairs & Center/Division Directors

Each Center Director, Division Director or other unit leader is responsible for providing adequate support to researchers in their administrative unit. They should ensure that adequate facilities are available to control biohazards, and to enable laboratory personnel to comply with applicable regulations and MBL policies.

2.3 Biological Safety Officer

The Environmental Health & Safety Manager serves as the designated Biosafety Officer (BSO) in accordance with NIH Guidelines. The BSO is responsible for providing guidance on the safe handling of biological agents and overall management of the biosafety program. The BSO is a member of the IBC.

The BSO's responsibilities include:

- Managing the biosafety program and implementation of IBC policies and procedures at the MBL.
- Ensuring records of infectious agents, recombinant DNA activities and Select Agents are kept current.

Biosafety Level 1 (BSL-1)

Biosafety Level 1 (BSL-1) is suitable for work involving well-characterized agents not known to often cause disease in healthy adult humans, and present minimal potential hazard to laboratory personnel and the environment. BSL-1 laboratories are not necessarily separated from the general traffic patterns in a building. Work is normally conducted on open bench tops using standard microbiological practices. Special containment equipment or facility design is not required but may be used as determined by appropriate risk assessment. Laboratory personnel must have specific training in the procedures conducted in the laboratory and must be supervised by a scientist with training in microbiology or a related scientific field.

Biosafety Level 2 (BSL-2)

Biosafety Level 2 (BSL-2) builds upon BSL-1. BSL-2 is suitable for work involving agents that pose moderate hazards to personnel and the environment. It differs from BSL-1 in that:

- Laboratory personnel have specific training in handling pathogenic agents and are supervised by scientists competent in handling infectious agents and associated procedures.
- Access to the laboratory is restricted when work is being conducted.
- All procedures in which infectious aerosols or splashes may be created are conducted in biological safety cabinets (BSCs) or other physical containment equipment.

Summary of Biosafety Levels for Infectious Agents.

Biosafety Level	Agent Characteristics
BSL-1	x Not known to consistently cause diseases in healthy adults.
BSL-2	x

3.2 Laboratory Specific Operating Procedures (SOPs)

A BSL-2 laboratory must adopt a laboratory-specific biosafety operating procedure that is tailored to the specific hazards and procedures associated with the laboratory. The required BSL-2 procedures must be described in the laboratory SOP with any additional practices that are to be implemented. This SOP should be submitted together with a research protocol to the IBC for review. A copy of the laboratory-specific SOP should be readily available and readily accessible in the laboratory (hard copy in a binder or in electronic form).

4 BIOLOGICAL RISK ASSESSMENT

Biological risk assessment is the process of evaluating biological research projects for potential hazards and characterizing the hazards into classes with proper containment and work practices to ensure research can be conducted safely and effectively. Biological risk assessment is an ongoing process during the biological research. Research involving recombinant DNA and synthetic nucleic acids will be reviewed by the MBL's IBC. Research involving animals will be reviewed by the MBL's Animal Care and Use Committee (IACUC).

4.1 Risk Assessment

The BSO, MBL Veterinarian, faculty and staff will participate in risk assessments. Information obtained from the risk assessment will be used as a guide for the selection of appropriate biosafety levels and microbiological practices, safety equipment, and facility design to limit personal exposure to biohazards.

The primary factors to consider in risk assessment and selection of safety precautions fall into two broad categories: (a) agent hazards and (b) procedure hazards.

During a risk assessment, the following factors should be evaluated:

(a) The biological and physical hazard characteristics of the agent.
(b) The sources likely to harbor the agent.
(c) Host susceptibility.
(d) The procedures that may disseminate the agent.
(e) The best method to effectively inactivate the agent.

The competence of the laboratory personnel to control hazards should be considered. This will depend on the training and technical proficiency, adhering to safe work practices by laboratory staff and the effectiveness of the containment equipment and facility safeguards.

The following five steps for conducting risk assessment are recommended:

(a) Identify the hazards of the agents and perform an initial assessment of risk.
(b) Identify the hazards of the experiment proced)

(e) Review the risk assessment with the PI/CDs, subject matter expert, BSO and the IBC, as appropriate.

4.2 Classification of Infectious Agents Based on Risk Groups

Biohazardous materials are classified according to risk levels. Following a risk assessment, biohazardous agents are placed into appropriate containment levels. Some biological materials that may not be normally considered as biohazardous may be regulated. There are four risk groups (RG) which are defined in Appendix B of the NIH Guidelines:

https://osp.od.nih.gov/biotechnology/nih-guidelines/

Microorganisms that are in RG1 require basic laboratory facilities and standard microbiological practices. The biological agents that are likely to be encountered in MBL research and teaching laboratories are RG1 or RG2 pathogens, designated as low and moderate hazard, respectively.

The progression from invasion to infection following contact with an infectious agent depends upon the route of transmission, infectious dose, pathogenicity of the agent, and resistance of the person exposed (innate or acquired immunity). Not all contacts result in infection. It is prudent, to assume virulence and handle biological agents at the appropriate biosafety level.

Classification of Infectious Microorganisms by Risk Group.

Risk Group	CDC/NIH/WHO	Examples
Risk Group 1	Agents not associated with disease in healthy adult humans. No or minimal individual and community risk.	Adeno-associated virus (AAV) Escherichia coli K-12 host vector systems
	Agents associated with human disease that is rarely serious and for which preventive/therapeutic interventions are often available. Moderate individual risk but low community risk.	Salmonella enterica Escherichia coli Staphylococcus aureus Toxoplasma Hepatitis B virus
	Agents associated with serious or lethal human disease for which preventive or therapeutic interventions may be available. High individual risk but low community risk.	Mycobacterium tuberculosis Bacillus anthracis West Nile virus Yellow fever HIV

RG3 and RG4 agents require more sophisticated engineering controls (e.g., facilities and equipment) than standard laboratories, and special handling and decontamination procedures. Laboratory research on RG3 and/or RG4 agents is not permissible at the MBL. Containment facilities and work practices for these agents have not been established.

4.3 Routes of Exposure

An infection occurs when disease-causing microorganisms enter the human body in sufficient numbers and by a particular route and overwhelm the body's defense system. Below are the routes of possible infection:

Ingestion

- Eating or drinking in the laboratory.
- Mouth pipetting.
- Transfer of microorganisms to mouth by contaminated fingers or articles.

Injection

- Accidental inoculation with a hypodermic needle or other sharp instrument.
- Abraded skin through cuts, scratches, skin rashes, etc.
- Animal bites or scratches.

Absorption

- Splashes of infectious material into the eye.
- Transfer of microorganisms to eyes by contaminated fingers.
- Through contact with skin.

Inhalation

- Inhalation of airborne microorganisms.

4.4 Exposure Sources

Exposure sources in the research laboratory are hazards that could result in infection of laboratory personnel and students or members of the public through work involving biological agents.

Some of the most common hazards that should be considered include the following:

4.4.1 Pathological Specimens

Any specimens from human and animal sources may contain infectious agents. Animals may harbor endogenous pathogens that are virulent for humans. When handling these animals or their tissues/body fluids, laboratory personnel should follow Universal Precautions.

The laboratory personnel are required to undergo Bloodborne Pathogen (BBP) training, in compliance with OSHA Bloodborne Pathogens Standard (29 CFR 1910.1030). The training will be provided by the BSO.

4.4.2 Laboratory Aerosols

Various laboratory procedures can generate biological aerosols that can be potentially hazardous to personnel and surrounding environment. Standard microbiological practices provide aerosol minimization techniques, but some special precautions can help enhance the level of prevention.

Standard Precautions:

- When working with infectious or potentially infectious materials, aerosol-generating procedures must be performed in a Biosafety Cabinet (BSC).
- Use respiratory protection whenever required (as approved by BSO).
- When a spill occurs, leave the area for 30 minutes to allow droplets to settle and aerosols to be removed by HVAC before cleaning up a spill.

Specific Precautions:

- **Pipetting:** Dispense liquids as close to the reservoir as possible. Rinse with the appropriate disinfectant before disposing pipettes or pipette tips. Take caution when removing pipette or pipette tips.
- **Proper Use of Syringes:** Avoid air bubbles. Do not recap needles of syringes unless using a safety device that permits singlehanded recapping.
- **Inoculation loops:** Use pre-sterilized, disposable loops when possible. When heat-sterilizing reusable loops, allow loop to cool before introducing it to any potentially infectious materials.
- **Decanting Liquids:** Pour carefully and as close to the vessel receiving as possible.
- **Animal inoculation:** Remove syringe needles from animals slowly, with a smooth and steady motion.
- **Vortexing:** Ensure all substances being vortexed are in a durable container with a tight-fitting lid.
- **-**

BSL-2	Risk Group 2 Agents	Biosafety Level 2
BSL-2	Human Derived Materials	Biosafety Level 2 OSHA Bloodborne Pathogens
BSL-2	None (Custodial, POM and Security)	Biosafety Awareness/Hazard Communication OSHA Bloodborne Pathogens
BSL-1 or BSL-2	Materials to be Autoclaved	Autoclave Training
BSL-1 or BSL-2	Shipping Hazardous Materials	Shipping Biological or Infectious Substances

6

- Inform the Biological Safety Officer at 508-289-7424 or safety@mbl.edu and your supervisor about the spill.

7 LABORATORY PRACTICES AND PROCEDURES

With the use of proper practices and procedures in the laboratory or work area, exposure to biohazards can be minimized. The following practices are important for the prevention of laboratory infection and disease, and for the reduction of the potential for contamination of experimental material.

7.1 Personal Hygiene

- Wash hands frequently after handling infectious materials.
- Always wash hands for at least 20 seconds before leaving the laboratory or work area.
- Do not touch your face with your hands.
- Use good microbiological techniques to protect your research, colleagues and yourself.
- Do not eat, drink, or store food, beverages or medications in the laboratory.
- Do not apply cosmetics, eye drops or lip balm in the laboratory.
- Do not handle contact lenses in the laboratory.

7.2 Laboratory Procedures for Handling Infectious Agents

- A laboratory specific biosafety manual must be developed outlining activities and Standard Operating Procedures (SOPs).
- Keep laboratory doors closed to restrict access to only authorized laboratory personnel.
- When RG2 (or higher) pathogens are used in long-term studies, post a biohazard sign at the laboratory entrance identifying the agents in use and dates for which used. In addition, all laboratories must have a door sign that states the names and contact numbers for emergency laboratory personnel.
- All laboratories must contain a sink for hand washing and access to

X

personnel in the proper use of PPE. PPE must not be taken home or worn outside the laboratory.

Appropriate laboratory PPE should be determined following a thorough risk assessment of the work being performed. For assistance in selecting PPE for work with biological materials, contact the BSO at 508-289-7424 or safety@mbl.edu.

The recommended laboratory PPE includes, but is not limited to, the following:

Laboratory Coats

- Coats are used to prevent contamination of the skin and street clothes. MBL offers a laboratory coat program

They are designed to provide three types of protection when used together with standard microbiological practices: (a) Personal protection from material inside the cabinet; (b) Protection for the material inside of the cabinet; and (c) Protection for the environment from the material inside of the cabinet.

8.1 High Efficiency Particulate Air (HEPA) Filters

Control of airborne particulate materials is achieved with HEPA filters that efficiently remove microscopic contaminants from the air. The HEPA filter removes particles H T X D O W R D Q G J U H D W H U W K D Q includes all bacteria, spores, and viruses with an efficiency of 99.99%.

Air flow is balanced so that some air is taken from the room and, together with sterile cabinet air, sucked into a horizontal grill at the front of the work surface. A BSC shall generally not be used for work with hazardous chemicals. The BSCs at MBL exhaust the contaminated air through HEPA filters back into the laboratory.

8.2 Classes of Biosafety Cabinets

Biological Safety Cabinet Characteristics

BSC CLASS, TYPE	Face Velocity (fpm)	Airflow pattern	Nonvolatile toxic chemicals /radionuclides	Volatile toxic chemicals or radionuclides
Class I	75	Enters at the front of cabinet. Exhausted through HEPA filter to the outside or room.	YES	YES (when exhausted outside)
Class II, A1	75	70% recirculated to the cabinet work area through HEPA filter. 30% balance can be exhausted through HEPA filter back into the room or to the outside through a canopy unit.	YES (small amounts)	

- Wipe work surface with an appropriate disinfectant (e.g. 70% ethanol).
- Plan your work and place everything needed for the procedure inside the BSC.
- Wipe items with disinfectant before placing inside the BSC.

During work:

- Avoid airflow disruption that could affect the level of protection provided by the BSC.
- Keep the BSC free of clutter.
- Do not place objects over the front air intake grille.
- Do not block the rear air intake grille.
- Limit traffic around the area when the BSC is in use.
- Ensure laboratory door is closed and avoid opening or closing the door if it is located near the BSC.
- Move arms slowly when removing or introducing items. Avoid sweeping arm motions.
- When working with infectious materials, change gloves when moving hands in and out of the cabinet.
- Keep all materials at least 4 inches inside the front sash.
- Keep clean materials at least one foot away from any aerosol generating activities to minimize the potential for cross contamination.
- Place any equipment that creates air turbulence in the back one third of the BSC and stop other work while the equipment is running.
- Do not operate a Bunsen burner inside the BSC to avoid disrupting air flow patterns within the BSC.
- Segregate contaminated and clean items. Work from “clean to dirty.”
- Clean up all spills in the cabinet immediately. Allow the BSC to run for about 3-5 minutes before resuming work.

After completion of work:

- Wash hands for 20s after completion of work.
- Using a new pair of gloves, wipe down all items with an appropriate disinfectant before removing from the hood.
- After removing all materials from hood, wipe all interior surfaces of the BSC with an appropriate disinfectant.
- Laboratory personnel should remove gloves and laboratory coats, and then wash hands for 20s as the final step in safe microbiological practices.
- Periodically decontaminate and clean under work grilles.

8.5 Certification of Biosafety Cabinets

The MBL requires that all BSCs installed on campus be tested and certified annually by technicians accredited by the National Science Foundation (NSF). BSCs must be tested and certified annually or if:

- A new cabinet is being installed.
- A cabinet has been moved.
- A cabinet requires maintenance and repairs.

All certification, maintenance, and repairs of BSCs are be conducted by an NSF-approved vendor contracted by the MBL.

9 APPROVAL FOR BIOHAZARD WORK AT MBL

All research at the MBL involving biohazards (or utilizing Non-Native Species) must be documented and submitted to the Institutional Biosafety Committee (IBC) for review and approval prior to the initiation of work. Approved protocols are valid for three years. For the purposes of the IBC, biohazards are defined as potentially infectious agents, organisms or materials, biological toxins, organisms containing recombinant DNA, or other genetically altered organisms and agents. Please note: if you plan on using or collecting biological agents, samples, etc. from live vertebrate animal sources or identifiable human sources, or your research will involve radiation/radioactive isotopes, you will must seek approval from the Institutional Animal Care and Use Committee (IACUC), Institutional Review Board (IRB), or Radiation Safety Committee (RSC).

The MBL's IBC shall meet at least once every six months (or as often as necessary) to discharge its institutional obligations. To register or amend research, download the application form "Institutional Biosafety Committee Biosafety Application" at:

https://www.mbl.edu/research/health-safety/biological-safety/biosafety-committee

9.1 Recombinant DNA

The National Institutes of Health (NIH) has developed guidelines on the use and containment of recombinant DNA materials in the research laboratory. Strict adherence to the NIH Guidelines for Research Involving Recombinant of Synthetic Nucleic Acid Molecules is critical to maintain effective regulatory oversight of rDNA at the MBL. The NIH Guidelines require individuals conducting such research to file a registration form with the IBC which must approve the protocols related to rDNA molecules.

The NIH Guidelines are designed to outline the practices for constructing and handling rDNA, both natural and synthetic, as well as cells, organisms, and viruses that contain these molecules. Recombinant DNA refers to either: (a) molecules constructed outside of living cells by joining natural or synthetic DNA segments to DNA molecule that can replicate in a living cell, or (b) molecules that result from the replication of those described in (a) above.

The NIH Guidelines specify different levels of approval and registration requirements (Sections III-A through Section III-F) that must be met prior to or upon initiation of work. Certain uses of rDNA (Section III-F) are exempt from any approval or registration requirements.

Summary of Experiments Falling Under NIH Guidelines

All recombinant DNA research proposals require the Responsible Researcher to make an initial determination of the required level of physical and biological containment. The NIH has developed six categories (**III-A** to **III-F**) addressing different types of rDNA research.

9.2 Research Involving Animals

The use of vertebrates in research is subject to federal law (Animal Welfare Act) and state regulations. The MBL complies with all federal and State and local regulations regarding animal use and care. All animal protocols using vertebrates or cephalopods must be reviewed and approved by the Institutional Animal Care and Use Committee (IACUC). To register or amend research involving animals, download the appropriate application form at:
https://www.mbl.edu/research/research-administration/compliance-and-ethics/iacuc

9.3 Select Agent Toxins

Some select biological toxins are highly regulated and subject to registration with the Centers for Disease Control and Prevention (CDC) or the United States Department of Agriculture (USDA), due to their potential to pose a severe threat to public health and safety. All users of select agents must complete an IBC application and must have appropriate training before commencement of work. Below are the exempt quantities for the select toxins. See MBL's Select Agent Toxin Policy:
https://www.mbl.edu/research/health-safety/biological-safety

Exempt Quantities of Select Agent Toxins

HHS Toxins [§73.3(d)(7)]	**Amount**
Abrin	1000 mg
Botulinum neurotoxins	1 mg
Short, paralytic alpha conotoxins	100 mg
Diacetoxyscirpenol (DAS)	10,000 mg
Ricin	1000 mg
Saxitoxin	500 mg
Staphylococcal Enterotoxins (Subtypes A, B, C, D, and E)	100 mg
T-2 toxin	10,000 mg
Tetrodotoxin	500 mg

10 DECONTAMINATION OF EQUIPMENT AND WORK AREAS

Any equipment that is going to be either moved for an area, reused or disposed must be decontaminated. This includes both single use and multiple use equipment.

Decontamination is defined as chemical or physical treatment that destroys microorganisms that produce disease to an acceptable level. This can be achieved through disinfection or sterilization. **Disinfection** is the physical or chemical treatment that destroys most microbial agents, except spores. Although viable microbes may still be present, they are below the level necessary to cause disease. **Stisi3Aszatl 0 Td[Ds**

organism from decontamination. For example, decontamination of a surface such as a

Hydrogen peroxide can be applied by spray or use of a damp towelette which is more appropriate near animal life.

For Hydrogen Peroxide to serve as an effective disinfectant, a longer contact time is required. In use of a low (0.25-0.50%) solution, the disinfected surface should be allowed to stay wet 5-10 minutes before wiping dry.

(d) Quaternary Ammonium Compounds

Less commonly used at MBL are solutions containing quaternary ammonium. These solutions are cationic detergents that have strong surface activity and are generally nontoxic (ideal for general use). They are active against Gram-positive bacteria and lipid-containing viruses, but less active against Gram-negative bacteria. They are ineffective against non-lipid-containing viruses. They are rendered inactivate by organic materials, anionic detergents or metal salts found in water.

11 BIOLOGICAL WASTE DISPOSAL

Proper handling and disposal of biohazardous waste is critical to prevent infection of personnel (laboratory workers, students, custodial staff, laboratory visitors, etc.) and release to the environment. The Commonwealth of Massachusetts regulations (105 CMR 480.000) and OSHA require that biohazardous waste be properly labeled, stored, and disposed of.

There are three categories of biohazardous waste: biohazardous solid waste, biohazardous sharps waste, and biohazardous liquid waste. At a minimum, all biohazardous waste must be labeled with the Universal Biohazard Symbol.

11.1 Biohazardous Solid Waste

Biohazardous solid waste includes microbial agents, tissue culture, and contaminated material (e.g., petri dishes, pipettes, contaminated glass, plastic plates, paper towels, gloves, etc.). Non-sharp biohazardous solid waste must be decontaminated before disposal using an autoclave.

Solid 20 (l)6 (abor)182Dd4 (er) 4.06 02 Ted(ar)-3ho4 (s)4-3 20 (t)2crdous Td[c3 (ps)4 (wdas)4 (

11.2 Biohazardous Sharps Waste

A sharp is any instrument that is capable of causing punctures, cuts, or scrapes to the body. Sharps includes but is not limited to needles, syringes, scalpels, razor blades, slides, coverslips, Pasteur pipettes, capillary tubes, sharp or broken glass, lancets, suture needles, and microtome blades. Use of sharps should be restricted to trained personnel and cases in which no alternative is available.

Sharps Precautions

- Avoid the use of needles and other sharps whenever possible. Use plastic ware instead of glassware if available.
- Minimize any contact with sharps by disposing of or storing them immediately after use.
- Needles must never be recapped, removed from the syringe, sheared, bent or broken. Dispose of the entire syringe-needle combination in a sharps container.
- Use a mechanical device to remove scalpel blades. Never use your fingers.
- Be careful during cleanup; some sharp items may be hidden in the waste materials.
- If a needle stick occurs, allow the wound to bleed for a few minutes, gently wash the area with water and soap, and then get medical attention immediately.

11.2.1 Disposal of Laboratory Sharps

Sharps are considered as regulated medical waste under federal and state regulations. They must not be disposed of as regular trash. All sharps must be disposed of into an approved, puncture resistant, sharps container.

- Sharps used with genetically modified and biological materials must be collected in a rigid, leak-proof, puncture resistant, mdy <<bl (mDC /edu /P <<

Disposal of uncontaminated laboratory glassware and broken glass should be p-placed into the standard broken “GLASS” cardboard boxes. However biologically contaminated glassware, place directly into a sharps containers for a small container (100ml or less). A larger contaminated container (>100ml) must be autoclaved prior to being placed into the broken glass box.

11.3 Biohazardous Liquid Waste

Biohazardous liquid waste is material that is contaminated with biological agents, including human blood and body fluids, liquid culture media, viral supernatant, and media from infected cells. Collect biohazardous liquid waste in closeable, rigid, plastic leak-proof containers labeled with the Universal Biohazard symbol. The waste must be decontaminated by autoclaving (see Section 10.1)

- The labeled mixed radioactive waste is picked up by the Radiation Safety Officer (RSO) and transported to the Radioactive Waste Storage Facility for packaging and

13.1 Import of Infectious Materials

The CDC's Import Permit Program (IPP) regulates the importation of infectious biological agents, infectious substances, and vectors of human disease into the United States. Items that most often require a permit from CDC are:

1. **Etiologic agents**:

13.4 Material Transport Agreements

Material Transfer Agreements (MTAs) address the exchange of research materials between individuals at separate organizations and address ownership, intellectual property, publications, and liability related to the research materials.

14 REGULATIONS AND GUIDELINES

1. CDC and the National Institute of Health (NIH) Guidelines on: Biosafety in